

Works with the TL-4015 Central Monitor



Cordless and Wireless System

IMPORTANT INFORMATION: Cordless and Wireless Systems rely on wireless technology which is subject to physical and environmental considerations. Please be aware that the transmitter will not be 100% accurate if it is out of range at any time. It is the end-user's responsibility to make sure that the product is used correctly and within range of each other. Cordless and wireless systems and devices are intended as an adjunct to good caregiving practices and are not a substitute for proper staffing and patient management practices. We recommend that all caregivers receive periodic training in the operation of these systems and that the devices are tested daily. It is important to read and follow the directions.

Read the instructions and Legal Disclaimer.

The Smart Caregiver Corporation devices are designed to be installed by the end user. As such, it is the entire responsibility of the buyer to insure that the system is properly installed and tested. Further, the system is not designed to replace good caregiving practices including, but not limited to:

- A.) Direct patient supervision
- B.) Adequate training for staff personnel for fall prevention and elopement
- C.) Testing of the system before each use

WARNING: The manufacturer does not claim that this device will stop elopement and/or stop falls. This device is designed to augment caregivers' comprehensive resident mobility management program. Failure to comply with the warning may result in injury or death. This device is not suitable for all individuals. Other devices may be required. This device is not a substitute for visual monitoring by a caregiver.

Where permissible, it is the seller's intent to limit any liability for special, consequential or incidental economic damages to refund of the purchase price or replacement of the goods. Seller disclaims any liability for compensatory or other damages arising out of the use of the goods.

Limited Warranty: Seller warrants that the goods sold will be free from defects of workmanship and manufacturing, for a warranty period from the date of sale. Other than this limited warranty, seller makes no express or implied warranties. Except for this limited warranty. No express or implied warranties are given and no affirmation of seller, by words or action will constitute a warranty. The goods being purchased are being sold on a "with all faults" basis. The entire risk as to the quality and performance of the goods is with the buyer. Should the goods prove defective following their purchase, the buyer, and not the manufacturer, distributor or retailer, assumes the entire cost of all necessary installation, servicing or repairs.



1229 North McDowell Blvd, Petaluma, CA 94954
csr@smartcaregiver.com | www.smartcaregiver.com
Toll Free: (800) 650-3637




IF YOUR SYSTEM IS PRE-PROGRAMMED, DO NOT FOLLOW THE STEPS BELOW

Instructions If Reprogramming Is Needed


Need Set-up Help? (800) 650-3637

Programming Call Button (4017-CB) & Reset Button (4017-RB) to Call Light (4017-ECA)


- 1**
- Insert 3-C batteries into the Call Light.
Re-Chargeable Batteries NOT Recommended
 - Set switch to High or Low position.



- 2**
- Insert pin (paper clip or toothpick) into program button on back of Call Light & hold until lights stay on solid then remove pin.




- 3**
- A.** Press Call Button until light comes on & then release.



The Call Light will start beeping.

- B.** Press Reset Button until light comes on & then release.



The Call Light will continue beeping.

- 4**
- A.** Place pin in program button on Call Light & quickly remove.
- B.** Call Light will do one long beep and lights will turn off.
- C.** Programming complete.
- D.** Test system.

*Silent volume setting should only be used when call buttons are programmed to the central monitor.



Need Set-up Help?


Call one of our friendly Customer Service Representatives at (800) 650-3637

Troubleshooting

To clear programming:
Monitor should be set to either "HIGH" or "LOW" on the left side of the monitor.

Insert pin (paper clip or toothpick) into program button on back of Call Light until lights stay on solid and then remove pin.

Insert pin again & hold. Wait for solid beep and lights turn off.



Remove pin.
Programming cleared.


Green Flashing Lights and Beeps =

- Check Power Source (Low Batteries)
- Check AC power connection
- (AC-04 Optional Component)



Green•Red•Green Flashing and Beeps =

- Signal Lost, Please Contact Smart Caregiver at number below for assistance
 - » Reprogram
 - » Buttons have 3-5 year battery



Call Smart Caregiver for assistance at (800) 650-3637